

**IN THE UNITED STATES
PATENT AND TRADEMARK OFFICE**

TITLE:

NURSE CALL INTERFACE AND METHOD OF OPERATION

INVENTORS:

**KEVIN W. KILE
ROBERT F. SANFORD
WILLIAM J. KANEWSKE, III**

BACKGROUND OF THE INVENTION

1. Field of The Invention

Applicant's invention relates to a nurse call interface and method of operation.

2. Background Information

The concept of notifying a nurse call station whenever a patient or resident exits the bed is not new; however, there are several different ways in which this concept can be carried out as is illustrated in the prior art.

U. S. Patent No. 4,020,482 issued to Feldl disclosed an elongated air bag that was placed under a bed mattress and connected to a pressure actuated electrical switch in a remote location. The switch of this invention is open as long as the patient's weight maintains an air pressure greater than the preset actuating pressure and closes when the patient's weight is removed lowering the pressure. When the pressure is lowered a signal is sent to a remote location. This invention incorporates a pair of pressure plates to increase the surface area sensing the weight change.

In the patent issued to Roberts, U.S. Patent No. 4,228,426 another monitoring system is disclosed which utilizes a connecting unit attached to a flat pressure activated switching means that is adapted for connection between a nurse call button and a communication panel in a hospital room. With this invention, a pair of electrical conductors indicate when the patient is in the bed. When the pair of electrical

1 conductors is pressed together, a signal is sent indicating the patient is still in the bed.

2 U.S. Patent No. 4,539,560 issued to Fleck et al discloses a bed departure
3 detection system that employs tape switch detectors in fixed positions in a bed
4 mattress. In the tape switch system, two resilient metal conductors pressed together
5 indicates the patient is still in bed. The detection system connects with existing nurse
6 call communication links to communicate signals to a nurse call station which indicate
7 when the patient has exited the bed.

8 In U.S. Patent No. 4,638,307 issued to Swartout, an apparatus is disclosed
9 that determines when weight has been lifted from a location. This apparatus
10 incorporates a sealed reed switch attached near the bottom of a compressible pad and
11 a magnet attached near the top of the pad. The reed switch is attached in such a way
12 that when the pad is compressed and the magnet is brought into close proximity to the
13 reed switch, the circuit is open and no alarm sounds, but when the weight of
14 compression is removed, the circuit closes and the alarm sounds.

15 U.S. Patent No. 5,353,012 issued to Barham et al discloses a detecting
16 apparatus that incorporates an elongated sensor for placement parallel to one axis of a
17 bed. The sensor has two conductive members, one being connected to a power source
18 wherein the weight of a body in the bed urges the conductive members together to
19 define an electrical path. Values are provided concerning position and activity levels
20 such that when these values are exceeded an alarm sounds.

21 For U.S. Patent No. 5,600,108 issued to Newham, an enclosure is disclosed for

1 mounting a bed monitoring system to a bed frame. This invention does include a
2 sensor mat for determining the presence of the patient.

3 U.S. Patent No. 5,844,488 issued to Musick provides for a sensor pad to
4 attaches to the top and across the width of a mattress having a central pressure switch
5 which indicates the presence of a patient in the center of the bed. An early warning
6 signal is attached to an edge switch which is provided to indicate when the patient has
7 moved from the center of the bed to the edge of the bed.

8 In U.S. Patent No. 6,025,782, the continuation in part U.S. Patent No.
9 6,297,738 issued to Newham, and US 2002/0070866 published June 13, 2002 for
10 Newham, a device for monitoring the presence of a person is disclosed using proximity
11 induced dielectric shift sensing. A control module is provided with a capacitive array
12 which supplies a suitable oscillator driver current and concurrently senses capacitance
13 value changes within the capacitive array induced through dielectric shifts within the
14 array brought about by the presence or absence of the patient's body. When the
15 capacitance changes in this system, a signal is sent to the nurse station.

16 For U.S. Patent No. 6,417,777 issued to Fitzgerald, et al, a patient monitoring
17 system is disclosed including a pressure sensitive mat that is sealed around its exterior
18 with the interior kept in communication with the atmosphere by way of flexible tubing
19 which encloses an electrical line.

20 U.S. Patent No. 6,544,200 issued to Smith et al discloses an electronic patient
21 monitor that incorporates a sensor system that automatically configures itself depending

1 on the sensor used and/or the environment in which the monitor is placed. This
2 monitor provides numerous controls that provide warning signals when the weight of
3 the patient is no longer detected in the bed.

4 Pending U.S. Application Publication No. US 2002/0067273 published June 6,
5 2002 discloses a patient monitor that includes a pressure sensor put under the patient.

6 This sensor provides a normally open switch having two contacts. The closing of these
7 contacts requires bending one of the contacts over a spacing means. In this manner
8 the two contacts are kept apart until the spacing means is forced shut. Pressure on the
9 pad over a preset amount arms the switch and release of the pressure causes an alarm
10 to sound.

11 The present invention is distinct from the prior art. The present invention is
12 intended to be plugged directly into the nurse call box on the wall in a hospital room.
13 The nurse call interface works in tandem with bed exit and chair exit sensor pads. In
14 this arrangement, if the patient or resident exits the bed or chair, the sensor pad sends
15 a signal to the present nurse call interface. The present nurse call interface in turn
16 converts the normally closed signal of the sensor pad to an opened signal that the nurse
17 call box can monitor.

18 A microprocessor is provided within the nurse call interface and is connected to
19 an LED, alarm and relay assembly. The relay assembly is connected to a nurse call
20 interface plug which plugs into a nurse call box wall plate. Microprocessor senses
21 voltage changes from sensor pad caused by weight change thereon. If a patient exits

1 the sensor pad, the voltage sensed by microprocessor exceeds 2.5 volts and
2 microprocessor sends a signal to the LED, alarm and relay assembly actuating each. A
3 nurse call button assembly can be provided as part of the nurse call interface system
4 which connects to the nurse call interface. If the button on this assembly is depressed,
5 the LED, alarm, and relay assembly are each actuated. When the relay assembly is
6 actuated, a signal is transmitted to the nurse call station.

7 **SUMMARY OF THE INVENTION**

8 It is an object of the present invention to provide a novel nurse call interface.

9 It is another object of the present invention to provide a novel nurse call interface
10 that is designed to be inserted directly into the nurse call box on the hospital room wall.

11 Still another object of the present invention is to provide a novel nurse call
12 interface that works in a tandem with a bed or chair sensor pad that senses a weight
13 change thereon.

14 An additional object of the present invention is to provide a novel nurse call
15 interface that is designed to notify the nurse station when the nurse call button is
16 depressed or the patient or resident exits a chair or bed where the sensor pad is
17 located.

18 Another object of the present invention is to provide a novel nurse call interface
19 that incorporates an opening that is used to accept signals from a sensor pad located in
20 a bed or chair to sense a weight change thereon.

21 Yet an additional object of the present invention is to provide a novel nurse call

1 interface that incorporates an opening that is used to accept signals from a nurse call
2 button.

3 Still another object of the present invention is to provide a novel nurse call
4 interface that incorporates a novel electrical assembly for additional nurse call signal.

5 It is another object of the present invention to provide a novel electrical assembly
6 for the nurse call interface that incorporates a voltage regulator, low voltage comparator,
7 microprocessor, LED, alarm, and relay assembly.

8 An additional object of the present invention is to provide a novel electrical
9 assembly that incorporates a voltage regulator that converts the battery voltage to a
10 steady voltage for use by the remainder of the nurse call interface.

11 Yet another object of the present invention is to provide a novel electrical
12 assembly with a low voltage comparator that recognizes when the battery has reached
13 the end of its life.

14 It is another object of the present invention to provide a novel electrical assembly
15 that has the ability to use existing code or accept new code to use in the system.

16 In satisfaction of these and related objectives, Applicant's present invention
17 provides a nurse call interface as part of a nurse call interface system which includes a
18 sensor pad to determine if a patient has left a chair or bed. Nurse call interface is
19 powered by a battery which is regulated by a voltage regulator. A microprocessor is
20 provided within the nurse call interface and is connected to an LED, alarm and relay
21 assembly. The relay assembly is connected to a nurse call interface plug which plugs

1 into a nurse call box wall plate. Microprocessor senses voltage changes from sensor
2 pad caused by weight change thereon. If a patient exits the sensor pad, the voltage
3 sensed by microprocessor exceeds 2.5 volts and microprocessor sends a signal to the
4 LED, alarm and relay assembly actuating each. A nurse call button assembly can be
5 provided as part of the nurse call interface system which connects to the nurse call
6 interface. If the button on this assembly is depressed, the LED, alarm, and relay
7 assembly are each actuated.

8 9 **BRIEF DESCRIPTION OF THE DRAWINGS**

10 Fig. 1 is a diagrammatic view of the preferred embodiment of the nurse call
11 interface system of the present invention.

12 Fig. 2 is a perspective view of the components of the preferred embodiment
13 of the nurse call interface system of the present invention.

14 Fig. 3 is a back view of the preferred embodiment of the nurse call interface
15 of the present invention.

16 Fig. 4 is a block diagram of the electrical schematic for the preferred
17 embodiment of the nurse call interface of the present invention.

18 Fig. 5 is a detailed electrical schematic of the preferred embodiment of the
19 nurse call interface of the present invention.

DETAILED DESCRIPTION OF THE PREFERRED EMBODIMENT

Referring to Fig. 1, a diagrammatic view of the preferred embodiment of the nurse call interface system 100 of the present invention is shown in relation to a nurse call box wall plate 102. A nurse call box wall plate 102 is provided that is standard in the industry. The nurse call box wall plate 102 is designed to be attached to the wall in a patient's or resident's room. Nurse call box wall plate 102 has a signaling lamp 104 that lights up when an alarm signal has been sent. When an alarm signal is sent to the nurse call box wall plate 102 from the nurse call interface 108 through nurse call interface plug 106, the circuit within the nurse call box wall plate 102 is closed. When the circuit to the nurse call box wall plate 102 is closed the signaling lamp 104 will light up inside the patient room. While the circuit to the nurse call box wall plate 102 remains open, signaling lamp 104 does not light up.

Nurse call box wall plate 102 has a nurse call interface opening 110 that is designed to accept nurse call interface plug 106 from nurse call interface 108. Nurse call box wall plate 102 may have additional openings 112 to accommodate additional plugs (not shown).

Shown adjacent to nurse call box wall plate 102 is nurse call interface 108 that incorporates nurse call interface plug 106 at its front. Nurse call interface plug 106 is a ¼ inch phone plug that recognizes a contact closure between the tip 106a and sleeve 106b of the plug, such a contact closure can be from a push button from a patient's bed or from a relay 250 (See Fig. 5). At the rear of nurse call interface 108 is a jack

1 114 (See Fig. 2) which is designed to accept a standard telephone connector plug or
2 RJ9 connector plug 116. While a RJ9 connector plug is specifically discussed in the
3 preferred embodiment, any appropriate sized connector and corresponding jack 114
4 (See Fig. 2) on the nurse call interface 108 can be used.

5 The RJ9 connector plug 116 is a four position, four conductor connector plug.
6 The RJ9 connector plug 116 is attached to a cable 118. Cable 118 carries a voltage of
7 5V. The cable 118 connects the RJ9 connector plug 116 to the sensor pad 120.

8 Fig. 2 is a perspective view of the components of the preferred embodiment of
9 the nurse call interface system 100 of the present invention. Nurse call interface 108
10 is provided that incorporates nurse call interface plug 106 at its front. At the rear of
11 nurse call interface 108 is jack 114 which is designed to accept a standard telephone
12 connector plug or RJ9 connector plug 116. Located next to jack 114 is opening 286
13 which is designed to accept a nurse call button system (not shown). Positioned next to
14 jack 114 is a LED light 122 to notify the user that the nurse call interface 108 is
15 turned on and working. In the preferred embodiment, the LED light 122 turns on for
16 200 milliseconds and turns off for 5 seconds; however, any flashing sequence could be
17 used. The in-use LED is green.

18 The RJ9 connector plug 116 is attached to a cable 118. Cable 118 carries a
19 voltage of 5V. The cable 118 connects the RJ9 connector plug 116 to the sensor pad
20 120 by way of a first part of a three pin Molex connector 124 on the cable 118 and a
21 corresponding mating part of the three pin Molex connector 126 on the sensor pad

1 120. The mating part of the three pin Molex connector 126 is connected to sensor pad
2 120 by way of an electrical strip 128. While a three pin Molex connector is specifically
3 discussed in the preferred embodiment, any appropriate sized connector and
4 corresponding opening on the sensor pad 120 can be used.

5 In Fig. 3 a back view of the preferred embodiment of the nurse call interface
6 108 of the present invention is shown. Nurse call interface 108 incorporates nurse
7 call interface plug 106 at its front and jack 114 at its rear. Located next to jack
8 114 is opening 286 which is designed to accept a nurse call button system (not
9 shown). Positioned next to jack 114 is a LED light 122. Nurse call interface 108
10 consists of a top portion 332 and a bottom portion 334. Housed within the top
11 portion 332 and bottom portion 334 is an electrical assembly (See Fig. 5) that is
12 responsible for the operation of the nurse call interface 108.

13 Fig. 4 is a block diagram of the electrical schematic of the electrical assembly
14 for the preferred embodiment of the nurse call interface 108. The electrical
15 assembly is powered by a 9V battery 200, but the voltage from the 9V battery 200
16 is regulated by voltage regulator 206. When the nurse call interface 108 is plugged
17 into nurse call box wall plate 102, the nurse call interface 108 is in standby mode.
18 When sensor pad 120 is plugged into the nurse call interface 108, the nurse call
19 interface 108 turns on. The nurse call interface 108 will sense a short circuit at pin
20 3 and pin 4 (See Fig. 5). Once the nurse call interface 108 powers up, the
21 microprocessor 212 sends a signal to LED 122 to indicate the power is on and the

1 nurse call interface 108 is working. Once the microprocessor 212 is powered on, it
2 senses voltage from the sensor pad 120. If this voltage rises above 2.5 volts, such
3 as when the patient or resident no longer puts weight on the sensor pad 120, the
4 microprocessor 212 sends a signal through relay 250 at step 1 to the nurse call
5 interface plug 106 where it is ultimately sent to the nurse call station. In addition,
6 the microprocessor 212 sends a signal (a) to LED 122 at step 2 to actuate LED 122
7 and (b) to alarm 278 at step 3 to actuate alarm 278. Alarm 278 has a sound level
8 of 75 db at 2 feet. When the sensor pad 120 is actuated by weight thereon, the
9 voltage sensed by microprocessor 212 is below 2.5 volts and no signal is sent to the
10 nurse call interface plug 106 because relay 250 is now not energized.

11 If the nurse call button 330 is depressed by the patient or resident, a signal
12 is sent to the alarm 278 to actuate alarm 278 and to the LED 122 to actuate LED
13 122. In addition, a signal is sent to the relay 250 which is then transmitted to the
14 nurse call interface plug 106 where it is ultimately passed to the nurse call station
15 (not shown). When nurse call button 330 is not depressed, no signal is sent.

16 The nurse call interface 108 defaults to a closed circuit if the 9V battery 200
17 drops below 5.8 volts. If the 9V battery 200 voltage drops below 5.8 volts, the low
18 voltage comparator 336 sends a signal to the microprocessor 212 which sends a
19 signal to the alarm 278 to actuate the alarm 278 and to the LED 122 to actuate the
20 LED 122. In addition, a signal is sent to the relay 250 which is transmitted to the
21 nurse call interface plug 106. This signal is ultimately sent to the nurse call station.

1 While the 9V battery 200 voltage remains above 5.8V, no signal is sent.

2 In Fig. 5 a detailed electrical schematic of the electrical assembly of the
3 preferred embodiment of the nurse call interface 108 of the present invention is
4 shown. Within nurse call interface 108 is a 9V battery 200 which provides power to
5 the nurse call interface 108. E1 contact 202 and E2 contact 204 have pigtail leads
6 which connect to 9V battery 200. Connected to 9V battery 200 is a voltage
7 regulator 206. Voltage regulator 206 is low power and converts the 9V from the 9V
8 battery 200 to a steady 5 volts which is used to supply the remaining components
9 of the nurse call interface 108. The output voltage from voltage regulator 206 next
10 passes through a low voltage comparator 336 which is grounded at 344. Positive
11 terminal of low voltage comparator 336 is connected to resistors 338 and 340
12 terminating in a common tie ground 346. Low voltage comparator 336 compares
13 the 9V battery 200 input to an internal reference voltage. When the voltage at the
14 low voltage comparator 336 falls below the reference voltage this information is
15 communicated to the microprocessor 212. Pin 4 of the low voltage comparator 336
16 is connected to pin 7 of the microprocessor 212 to communicate the information
17 that the 9V battery 200 has reached the end of its life and should be replaced.
18 Resistor 342 is connected between voltage comparator 336 and microprocessor
19 212. When the 9V battery 200 has reached this predefined lower voltage level, the
20 LED light 122 flashes and an alarm 278 sounds. In addition, a signal is sent to
21 relay 250 which is transmitted to nurse call interface plug 106. While 9V battery

200 voltage remains above 5.8 volts, no signal is sent to nurse call interface plug 106.

Value and Maximum Tolerance of Component Parts

Part #	Value	Maximum Tolerance
218	140K	1%
220	100K	5%
240	0.1 μ F50V	10%
252	330 μ H	10%
254	330 μ H	10%
312	4.7K	5%
314	1K	5%
280	4.7K	5%
282	1K	5%
300	10K	5%
316	0.1 μ F50V	10%
302	1M	5%
306	1K	5%
338	100K	5%
340	100K	5%
342	10K	5%

Pin 1 carries the 5 volts from voltage regulator 206 out to the remainder of the nurse call interface 108. Resistors 218 and 220 set the low voltage set point for the low voltage comparator 336. Pins 2, 3, 6 and 7 are grounded for proper performance.

For the microprocessor 212, pin 1, designated VDD, supplies 5 volts to the microprocessor 212. Pin 8, designated Vss, leads to a common tie ground 238. Capacitor 240 is a standard bypass capacitor which provides a low impedance AC path

1 across the circuit to prevent the power from the microprocessor 212 from propagating
2 into the remainder of the electronics for the nurse call interface 108. Pin 2 of the
3 microprocessor 212 is an output from a timer built into the microprocessor 212. The
4 output from pin 2 goes through steering diode 244 to the LED 122. The steering diode
5 244 prevents the nurse call interface 108 from discharging through the LED 122. If
6 there is current flowing through steering diode 244, i.e. pin 2 goes high, then LED 122
7 lights up. Resistor 306 is a current limiting resistor for LED 122 which is connected to
8 a common tie ground 308.

9 Pin 3 of the microprocessor 212 is an output that turns on transistor 248, which
10 causes current flow through coil 322 that closes switch 251 of relay 250. Relay 250
11 is a normally open, single-pole, single-throw relay. Diode 324 allows for discharge of
12 coil 322 when opening relay switch 251. When a patient or resident exits the sensor
13 pad 120, depresses a nurse call button 330, or the 9V battery 200 drops to a
14 predefined voltage level, relay switch 251 of relay 250 is closed. When switch 251 is
15 closed, contacts 256 and 258 are connected together. These contacts 256 and 258
16 are located within the nurse call interface plug 106. When nurse call interface plug
17 106 is plugged into nurse call box wall plate 102 (See Fig. 1), the current through
18 contacts 256 and 258 is transmitted to the nurse call box wall plate 102 (See Fig. 1).
19 Inductors 252 and 254 located along the line are used to control electrostatic
20 discharges (ESDs). Transistor 248 has resistors 312 and 314 connected to the base
21 which are necessary to limit the base current in transistor 248.

1 Pin 4 of microprocessor 212 is part of the programming system for
2 microprocessor 212. Pin 4 of microprocessor 212 receives programming code
3 information from pin 1 of connector 262. Connector 262 has the ability to use existing
4 code or accept new code to use in the system. Pin 1, designated VPP, is the power
5 supply for the program. Pin 4 of connector 262, designated ICSPC, is an in-circuit
6 serial programming clock input. Pin 3 of connector 262, designated ICSPD, is an in-
7 circuit serial programming data I/O. Pin 2 of connector 262, designated GND, is a
8 ground leading to a common tie ground 272.

9 Pin 5 of microprocessor 212 is an output which connects to the base of
10 transistor 276. Transistor 276 has two base resistors 280 and 282 to limit the
11 base current of the transistor 276. Transistor 276 is connected to alarm 278 and
12 can turn alarm 278 on which emits an audible tone. The current to alarm 278 is
13 amplified by transistor 276.

14 Pin 6 of microprocessor 212 is responsible for receiving input from sensor
15 pad 120 through pin 1 of jack 114. Lead 288 connects 9V battery 200 at 204 to
16 pin 4 of jack 114. When sensor pad 120 is plugged into jack 114 via RJ9
17 connector plug 116, the nurse call interface 108 automatically turns on. Pin 3 of
18 jack 114 connects 9V battery 200 to a common tie ground 294. There is a short
19 between pin 3 and pin 4 of jack 114 to ground 294. When this short is
20 transmitted, the nurse interface system 108 begins operating. Pin 1 and pin 2 of

1 jack 114 sense a reduction in resistance whenever the sensor pad 120 is depressed
2 by current flow through resistor 300, which is connected to input pin 6 of
3 microprocessor 212. Resistor 300 provides isolation for pin 6 of microprocessor
4 212 when new code is being loaded into connector 262. Capacitor 316 smooths
5 out the signals received at pin 1 of jack 114.

6 Referring to Fig. 4, once the nurse call interface 108 powers up, the
7 microprocessor 212 sends a signal to LED 122 to indicate the power is on and the
8 nurse call interface 108 is working. Once the microprocessor 212 is powered on, it
9 senses voltage from the sensor pad 120. If this voltage rises above 2.5 volts, such
10 as when the patient or resident removes their weight from the sensor pad 120, the
11 microprocessor 212 sends a signal through relay 250 to the nurse call interface plug
12 106 where it is ultimately sent to the nurse call station. In addition, the
13 microprocessor 212 sends a signal to LED 122 to actuate LED 122 and alarm 278
14 to actuate alarm 278. When the patient's weight is on the sensor pad 120, the
15 voltage sensed by microprocessor 212 will be below 2.5 volts and no signal is sent
16 to the nurse call interface plug 106.

17 The patient or resident typically has a nurse call button system (not shown) in
18 the bed with him or her. The nurse call button system (not shown) typically consists of
19 a nurse call button 330 and a cable (not shown) with the cable (not shown) being
20 attached to the nurse call box wall plate 102. That nurse call button 330 could be
21 plugged into opening 286 of Fig. 2 or in the additional opening 112 (see Fig. 1).

1 When the nurse call button 330 is depressed by the patient or resident, current flows
2 through the coil 322 of relay 250 causing relay switch 251 to close. In this manner,
3 the patient has the ability to call the nurse without removing weight on the sensor pad
4 120. In addition, when the nurse call button 330 is depressed by the patient or
5 resident, steering diode 304 causes alarm 278 to turn on and emit an audible tone.
6 Further, the current from opening 286 goes through steering diode 310 to LED 122.
7 The steering diode 310 prevents the nurse call interface 108 from discharging through
8 the LED 122. If there is current flowing through steering diode 310 then LED 122
9 lights up. When the nurse call button 330 is not depressed, no signal is sent through
10 opening 286.

11 Although the invention has been described with reference to specific
12 embodiments, this description is not meant to be construed in a limited sense. Various
13 modifications of the disclosed embodiments, as well as alternative embodiments of the
14 inventions will become apparent to persons skilled in the art upon the reference to the
15 description of the invention. It is, therefore, contemplated that the appended claims
16 will cover such modifications that fall within the scope of the invention.